


Air Sampling Results (blue dots)



* The data gap for the date 01/04/2009 is due to the postponement of particulate monitoring based on humidity (suspended aerosol) interference with the instrument.

* The particulate monitor is affected by humidity. High relative humidity causes the instrument to read higher than the true particulate level in air. Air monitoring for particulate was temporarily suspended for several hours on 01/03/09 and 01/04/2009 due to very high relative humidity levels